



**MANUFACTURER:**

Globalcare Medical Technology Co., Ltd.  
7th Building, 39 Middle Industrial Main Road, European Industrial Zone,  
Xiaolan Town, 528415 Zhongshan City, Guangdong Province,  
PEOPLE'S REPUBLIC OF CHINA

**Product Category:**

Blood Pressure Measuring Equipment

**Product Code:**

GCE604

**Product Description:**

Veroval® Upper Arm Blood Pressure Monitor

**Classification - Annex IX:**

Class IIa, Rule 10

**Conformity Assessment Route:**

ANNEX VII (Internal Production Control) in conjunction with  
ANNEX V (QA Production) of CD 93/42 EEC(MDD)

We, the manufacturer, herewith declare that the stated medical devices meet the transposition into national law, the provisions of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices; including, at 21 march 2010, the amendments by Council Directive 2007/47/EEC.

All supporting documentation is retained at the premises of the manufacturer.

Globalcare Medical Technology CO., LTD. is exclusively responsible for this EC Declaration of Conformity.

**Standards applied:** See attached list

**Notified Body:**

TÜV SÜD Product service GmbH  
Ridlerstr 65, D-80339 München, Germany

**Identification Number**

0123

**(EC) Certificate(s):**

G2 088855 0011 Rev.00



**European Representative:**

Donawa Lifescience Consulting Srl  
Piazza Albania, 10  
00153 Rome  
Italy

**Place, Date of Declaration:**

Zhongshan, 2020-04-02

**Signature:**

Name: Lambert Zhao  
Position: General Manager